

Product Code

CNG050S290NCAAEZZZ

PUCK700

Serial ID

15333411

34283811

Order ID

10500174

Line Item

3.1 1

Customer Tag

SN23011361 51



Process

Process ID : 1.36932292

Process Time : 2022.09.06 7:50:49

Process Stand : TSM1C@SSCB:1

Stand Uncertainty : +/-0.030%

Fluid : H2O

100% Rate : 38.6 KG/MIN

Pickoff : 1

Max Rate P/T : 27.41 PSIG/22.2 C

Detail



Results

Status : PASS

D1 : 0

D2 : 1

K1 : 4088.735

K2 : 4258.664

DT : 4.25

FD : 0

DTG : 0

DFQ1 : 0

DFQ2 : 0

FlowCal : 139.114.50

FFQ : 0

FTG : 0

DensCal : 04089042594.25

FCF : 139.11


FT : 4.5

*Neil Wilson*

WILSON, NEIL

Technician

This certificate is produced by an electronic data system and is valid without signature.



Flow (%)	Flow Rate (kg/min)	Meter Total (kg)	Reference Total (kg)	Error (%)	Specification (±%)
100.0	38.6	39.51932	39.52069	-0.003	0.500
10.0	3.86	5.616815	5.619313	-0.044	0.500
50.0	19.3	19.18356	19.18308	0.003	0.500
100.0	38.6	39.26456	39.26558	-0.003	0.500

Product Code

CNG050S290NCAAEZZZ

2700I13ABAEZMZ

PUCK700

Serial ID

15333411

12219366

34283811

Order ID

10500174

10500174

Line

3.1

3.33

Customer Tag

SN 23011361 51

Process

Process ID : 1.36932850

Process Time : 2022.09.06 12:08:19

Process Stand : CONFIGURATION@SSCB

Sensor

D1 : 0

D2 : 1

DFQ1 : 0

DFQ2 : 0

DT : 4.25

DTG : 0

Dens PCF : 0

Density Meter Factor : 1

FCF : 139.11

FD : 0

FFQ : 0

FT : 4.5

FTG : 0

Flow PCP : 30

Flow PCF : 0

K1 : 4088.735

K2 : 4258.664

Mass Flow Meter Factor : 1

Volume Flow Meter Factor : 1

Units

Density Unit : G/CM3

GSV Flow Unit : SCFM

Mass Flow Unit : G/SEC

Pressure Unit : POUNDS/SQUARE INCH

Special GSV Base Time Unit : MIN

Special GSV Base Volume Unit : Standard cubic feet

Special GSV Conv Factor : 1

Special GSV Flow Unit Text : NONE

Special GSV Total Text : NONE

Special Mass Base Unit : G

Special Mass Conv Factor : 1

Special Mass Flow Text : NONE

Units

Special Mass Time Unit : SEC

Special Mass Total Text : NONE

Special Volume Base Unit : L

Special Volume Conv Factor : 1

Special Volume Flow Text : NONE

Special Volume Time Unit : SEC

Special Volume Total Text : NONE

Temperature Unit : C

Volume Flow Unit : L/MIN

MVD Channel Assignments

Channel B Power : Active (internally powered)

Assignments

Event 1 Variable : Density

Event 2 Variable : Density

Frequency1 Scaling Method : Frequency = Flow

Frequency Variable 1 : Mass Flow Rate

mA1 Variable : Mass Flow Rate

Ranges

Event 1 Setpoint : 0

Event 1 Type : Event Low (Event "OFF" if PV > SP)

Event 2 Setpoint : 0

Event 2 Type : Event Low (Event "OFF" if PV > SP)

Frequency1 Active State : Active High

Frequency1 Hertz : 1000

Frequency1 Output Mode : Single

Frequency1 Pulses/Unit : 1.554404

Frequency1 Rate : 643.3333

Frequency1 Units/Pulse : 0.64333333

mA1 LRV : 0

mA1 URV : 643.3333

Faults

Frequency1 Fault Behavior : Upscale

Faults

SN 23011901 51

Frequency1 Fault Value : 15000  
mA1 Fault Behavior : Downscale (Default)  
mA1 Fault Value : 2

Other

Calibration Process ID : 1.36932292  
Core Software Rev : 35  
Density Cutoff : 0.2  
Density Damping : 0.8  
Density High Limit : 5  
Density Low Limit : 0  
Direction : FORWARD  
Fault Dwell Time : 0  
Feature Key : 1  
Flow Damping : 0.8  
HART Device ID : 4261421  
LD Type : 0  
Mass Flow Cutoff : 1.836  
Pressure Comp Line Pressure : 0  
Pressure Compensation State : OFF  
RS485 Baud : 1200 baud  
RS485 Parity : Odd  
RS485 Protocol : HART  
Slug Duration : 0  
Tag :  
Temperature Damping : 2.4  
Transmitter Software Rev : 80  
Volume Flow Cutoff : 0.11016



**Tulsa Gas Technologies, Inc.**  
4809 S. 101<sup>st</sup> East Ave Tulsa, OK 74146  
PHONE: 918-665-2641 FAX: 918-665-2657

1/12/2023

Dispenser Serial Number 23011361 (H1)

## Micro Motion Transmitter Configuration

Required settings for correct operation of Micro Motion mass flow meter.

Transmitter Model Number: 2700  
Sensor Model Number: CNG050  
Transmitter Serial Number: 12219366  
Sensor Serial Number: 15333411  
Flow Calibration Factor: 139.114.50  
Flow Units: LB/min

### Communication on RS-485

Protocol: Modbus ASCII 7 Bit  
Modbus Address: 1  
Baud Rate: 9600  
Parity: Even  
Stop Bits: 1

### HART Communication

Superimposed on Primary mA (PV)

### Analog Output (4-20 mA)

Analog Variable (PV): Mass Flow  
Lower Range Value: 0.2500 lb/min  
Upper Range Value: 300.000 lb/min  
mA Cutoff: 0.0000 lb/min

### Freq/Rate

Frequency variable (TV): Mass Flow  
Frequency Cutoff: 0.2500 lb/min  
Pulses per Unit: 1000.00000 per lb

### Temperature

Temp Units: deg F

Product Code

CNG050S290NCAAEZZZ

PUCK700

Serial ID

15335180

34283617

Order ID

10500174

Line Item

3.1 2

Customer Tag

SN 23011361 SZ



Process

Detail



Process ID : 1.36931201

Process Time : 2022.09.03 0:00:21

Process Stand : TSM1C@SSCB:1

Stand Uncertainty : +/-0.030%

Fluid : H2O

100% Rate : 38.6 KG/MIN

Pickoff : 1

Max Rate P/T : 29.93 PSIG/22.3 C

Results

Status : PASS

D1 : 0

D2 : 1

K1 : 4069.268

K2 : 4237.435

DT : 4.25

FD : 0

DTG : 0

DFQ1 : 0

DFQ2 : 0

FlowCal : 139.724.50


FFQ : 0

FTG : 0

DensCal : 04069042374.25

FCF : 139.72

FT : 4.5



Flow (%)	Flow Rate (kg/min)	Meter Total (kg)	Reference Total (kg)	Error (%)	Specification (±%)
100.0	38.6	39.54827	39.55923	-0.028	0.500
10.0	3.86	5.616107	5.61407	0.036	0.500
50.0	19.3	19.12341	19.12299	0.002	0.500
100.0	38.6	39.39066	39.39795	-0.019	0.500

  
 DINESH KARKI  
 Technician

This certificate is produced by an electronic data system and is valid without signature.

Product Code

CNG050S290NCAAEZZZ

2700I13ABAENZWZ

PUCK700

Serial ID

15335180

12219365

34283617

Order ID

10500174

10500174

Line Item

3.1

3.33

Customer Tag

5N23011361 SL

Process

Process ID : 1.36932785

Process Time : 2022.09.06 11:21:33

Process Stand : CONFIGURATION@SSCB

Sensor

D1 : 0

D2 : 1

DFQ1 : 0

DFQ2 : 0

DT : 4.25

DTG : 0

Dens PCF : 0

Density Meter Factor : 1

FCF : 139.72

FD : 0

FFQ : 0

FT : 4.5

FTG : 0

Flow PCP : 30

Flow PCF : 0

K1 : 4069.268

K2 : 4237.435

Mass Flow Meter Factor : 1

Volume Flow Meter Factor : 1

Units

Density Unit : G/CM3

GSV Flow Unit : SCFM

Mass Flow Unit : G/SEC

Pressure Unit : POUNDS/SQUARE INCH

Special GSV Base Time Unit : MIN

Special GSV Base Volume Unit : Standard cubic feet

Special GSV Conv Factor : 1

Special GSV Flow Unit Text : NONE

Special GSV Total Text : NONE

Special Mass Base Unit : G

Special Mass Conv Factor : 1

Special Mass Flow Text : NONE

Units

Special Mass Time Unit : SEC

Special Mass Total Text : NONE

Special Volume Base Unit : L

Special Volume Conv Factor : 1

Special Volume Flow Text : NONE

Special Volume Time Unit : SEC

Special Volume Total Text : NONE

Temperature Unit : C

Volume Flow Unit : L/MIN

MVD Channel Assignments

Channel B Power : Active (internally powered)

Assignments

Event 1 Variable : Density

Event 2 Variable : Density

Frequency1 Scaling Method : Frequency = Flow

Frequency Variable 1 : Mass Flow Rate

mA1 Variable : Mass Flow Rate

Ranges

Event 1 Setpoint : 0

Event 1 Type : Event Low (Event "OFF" if PV > SP)

Event 2 Setpoint : 0

Event 2 Type : Event Low (Event "OFF" if PV > SP)

Frequency1 Active State : Active High

Frequency1 Hertz : 1000

Frequency1 Output Mode : Single

Frequency1 Pulses/Unit : 1.554404

Frequency1 Rate : 643.3333

Frequency1 Units/Pulse : 0.64333333

mA1 LRV : 0

mA1 URV : 643.3333

Faults

Frequency1 Fault Behavior : Upscale

Faults

Frequency1 Fault Value : 15000  
mA1 Fault Behavior : Downscale (Default)  
mA1 Fault Value : 2

Other

Calibration Process ID : 1.36931201  
Core Software Rev : 35  
Density Cutoff : 0.2  
Density Damping : 0.8  
Density High Limit : 5  
Density Low Limit : 0  
Direction : FORWARD  
Fault Dwell Time : 0  
Feature Key : 1  
Flow Damping : 0.8  
HART Device ID : 4261415  
LD Type : 0  
Mass Flow Cutoff : 1.836  
Pressure Comp Line Pressure : 0  
Pressure Compensation State : OFF  
RS485 Baud : 1200 baud  
RS485 Parity : Odd  
RS485 Protocol : HART  
Slug Duration : 0  
Tag :  
Temperature Damping : 2.4  
Transmitter Software Rev : 80  
Volume Flow Cutoff : 0.11016

SN 23011361 S2



**Tulsa Gas Technologies, Inc.**  
4809 S. 101<sup>st</sup> East Ave Tulsa, OK 74146  
PHONE: 918-665-2641 FAX: 918-665-2657

1/12/2023

Dispenser Serial Number 23011361 (H2)

## Micro Motion Transmitter Configuration

Required settings for correct operation of Micro Motion mass flow meter.

Transmitter Model Number: 2700  
Sensor Model Number: CNG050  
Transmitter Serial Number: 12219365  
Sensor Serial Number 15335180  
Flow Calibration Factor: 139.724.50  
Flow Units LB/min

### Communication on RS-485

Protocol Modbus ASCII 7 Bit  
Modbus Address 1  
Baud Rate 9600  
Parity Even  
Stop Bits 1

### HART Communication

Superimposed on Primary mA (PV)

### Analog Output (4-20 mA)

Analog Variable (PV) Mass Flow  
Lower Range Value 0.2500 lb/min  
Upper Range Value 300.000 lb/min  
mA Cutoff 0.0000 lb/min

### Freq/Rate

Frequency variable (TV) Mass Flow  
Frequency Cutoff 0.2500 lb/min  
Pulses per Unit 1000.00000 per lb

### Temperature

Temp Units deg F